

TECHNICAL DATA

Fluke BP291 4800 mAh High Capacity Li-Ion Battery



Key features

- 4800 mAh rechargeable Li-ion battery
- Compatible with ScopeMeter® 190 Series III, 190 Series II, MDA-500 Series Motor Drive Analyzers and 430-II Series Power Quality Analyzers

Product overview: Fluke BP291 4800 mAh High Capacity Li-Ion Battery

The Fluke BP291 Li-ion Battery is a 4800 mAh high-capacity battery for use in Fluke 190 Series II and Series III ScopeMeter Test Tools, MDA-500 Series III Motor Drive Analyzers and 430-II Series Power Quality Analyzers. BP291 High-Capacity Li-ion batteries include a charge-status indicator so you can quickly and easily verify charge status in the field, and may be charged either in the instrument itself, or using the option [EBC290 External Charger](#).

Specifications: Fluke BP291 4800 mAh High Capacity Li-Ion Battery

Specifications*	
Battery	Li-ion, 10.8 V, 5000 mAh, 54 Wh configuration 3ICR19/66-2
Safety	Pollution degree 2

Operating temperature	0 °C to +50 °C (charging may slow or stop at environmental temperatures >40 °C)
Storage temperature	-20 °C to +60 °C
Recommended storage temperature	<35 °C
Humidity	0 °C to 10 °C noncondensing 10 °C to 30 °C @ 95 % (±5 %) rh 30 °C to 40 °C @ 75 % (±5 %) rh
Altitude	3000 m
Weight	330 g
Dimensions	22 mm x 150 mm x 59 mm

Ordering information



Fluke BP291

4800 mAh High capacity Li-ion battery for Fluke 190 seriesII

Fluke. *Keeping your world up and running.®*

Fluke Europe B.V.
P.O. Box 1186
5602 BD Eindhoven
The Netherlands
www.fluke.com/en

©2023 Fluke Corporation. All rights reserved.
Data subject to alteration without notice.
07/2023

For more information call:
In Middle East/Africa
+31 (0)40 267 5100

**Modification of this document is not permitted
without written permission from Fluke Corporation.**